

ACUMED™

K963118

*Quality Orthopaedic Instruments and Implants***Enclosure D - 510(k) Summary**

OCT 10 1996

This summary regarding 510(k) safety and effectiveness information is being submitted in accordance with the requirements of SMDA 1990 and 21 CFR 807.93.

The 2.7mm Extremity Bone Screw is intended for fracture and osteotomy fixation of the upper and lower extremities. This screw is not intended for usage in the spine and is not for use with any available bone plates, washers, etc. This screw is available in lengths of 12mm, 14mm, 16mm, 18mm, 20mm, 22mm, and 24mm. The screw is driven by a conventional 2.5mm hex head. Hex dimensions are in accordance with ASTM F 543. The Extremity Screw is manufactured from a titanium alloy per ASTM F 136 and is provided non-sterile. On file at Acumed is test data which shows that the screw can be successfully steam sterilized under specific process parameters which will obtain a resulting SAL of 10^{-6} . As the screw is provided non-sterile, packaging information is not provided.

The 2.7mm Extremity Bone Screw is similar to Synthes' 3mm Cannulated Screw System in design, function, indications, and surgical technique and is expected to perform as well as similar devices.



OCT 10 1996

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Ms. Shari L. Jeffers
Quality Regulatory Coordinator
Acumed, Inc.
10950 Southwest 5th Street, Suite 170
Beaverton, Oregon 97005

Re: K963118
Extremity Bone Screw
Regulatory Class: II
Product Code: HWC
Dated: August 8, 1996
Received: August 12, 1996

Dear Ms. Jeffers:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act.

The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Note that labeling or otherwise promoting a device for pedicular screw fixation/attachment would cause the device to be adulterated under 501(f)(1) of the Act. This device, if intended for use in pedicular screw fixation/attachment, would be found not substantially equivalent and would be a class III device under Section 513(f) of the Act. Class III devices are required to have an approved premarket approval (PMA) application prior to marketing. Accordingly:


1. All labeling for this device, including the package label, must state that there are labeling limitations. The package insert must prominently state that the device is intended for the specific use(s) described in the enclosure only; and
2. You may not label or in any way promote this device for pedicular screw attachment to, or fixation of the cervical, thoracic or lumbar vertebral column. If this device is a screw with outer diameters of 3 mm - 10 mm and overall lengths of 10 mm - 75 mm inclusively, the labeling must include the following statement, "**WARNING:** This device is not approved for screw attachment or fixation to the posterior elements (pedicles) of the cervical, thoracic, or lumbar spine." Any pedicular screw fixation/attachment to the cervical, thoracic or lumbar vertebral column of this device is considered investigational and may only be investigated as a significant risk device in accordance with the investigational device exemption (IDE) regulations under 21 CFR, Part 812. All users of the device for pedicular screw fixation/attachment must receive approval from their respective institutional review boards (IRBs) and the Food and Drug Administration (FDA) prior to conduct of the investigation.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the Good Manufacturing Practice for Medical Devices: General (GMP) regulation (21 CFR Part 820) and that, through periodic GMP inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal Laws or Regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification immediately. An FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or at (301) 443-6597.

Sincerely yours,



Celia M. Witten, Ph.D., M.D.
Director
Division of General and
Restorative Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

Page 1 of 1

510(k) Number (if known): K963118

Device Name: Extremity Bone Screw

Indications for Use: Fracture and osteotomy fixation of the upper and lower extremities.

(Division Sign-Off)
Division of ~~General Restorative~~ Devices
510(k) Number K963118

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

Prescription Use
(Per 21 CFR 201.109)

OR

Over-The-Counter Use _____

(Optional Format 1-2-96)